

File :Q:\GCMS-04 Minimal\DATA\053119\07-D
Operator : RFXM/KM
Acquired : 31 May 2019 19:00 using AcqMethod VOC-SCAN 24 DECKER.M
Instrument : GCMS4
Sample Name: M-554-R1
Misc Info : RR
Vial Number: 60



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

