

File : L:\QC Data\GCMS-09 Minimal\DATA\10202022\30 -SAK.D
Operator :
Acquired : 20 Oct 2022 21:47 using AcqMethod Hy VOC 24 AMU DB-5MS UI20M 0.18 50 split .M
Instrument : GCMS 9
Sample Name: M-602-SS-100X
Misc Info : a,a,a-Trifluorotoluene in MeOH
Vial Number: 132



Column: DB-5MS UI, 20m x 0.18mm ID x 0.18µm
Oven Program: 45°C, 13°C/min to 290°C (hold 5min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector temp 270°C

