

File :D:\MassHunter\GCMS\1\data\20221227\02 sak.D
Operator : Organic QC Lab
Acquired : 27 Dec 2022 08:03 using AcqMethod Hy_Pest_15cm_S10.M
Instrument : GCMS 8
Sample Name: M-608.2
Misc Info :
Vial Number: 148



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

