

File :Q:\GCMS-08 Minimal\DATA\20191104\15 -.D
Operator : KM / JP / ZA
Acquired : 04 Nov 2019 15:27 using AcqMethod KERLA-2017_S100.M
Instrument : GCMS 8
Sample Name: M-625-03-10X
Misc Info : Benz(a)anthracene-d12 in Dichlorom
Vial Number: 131



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

