

File :D:\MassHunter\GCMS\1\DATA\08022101\06-ALR.D
Operator : Organic QC department
Acquired : 02 Aug 2021 12:59 using AcqMethod KERLA-2017_Short_15Min.M
Instrument : 5977B MSD
Sample Name: M-625-06-10X
Misc Info : 4,4'-Dibromooctafluorobiphenyl in DCM
Vial Number: 13



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

