

File :L:\QC Data\GCMS-07 Minimal\DATA\03212201\03 - SAK.D
Operator : Organic QC department
Acquired : 21 Mar 2022 17:40 using AcqMethod KERLA-2017.M
Instrument : 5977B MSD
Sample Name: M-625-BN-1
Misc Info :
Vial Number: 127



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

