



=====

Acq. Operator : SYSTEM	Seq. Line : 2
Acq. Instrument : LCMS-1	Location : P1-C4
Injection Date : 1/3/2024 11:11:48 AM	Inj : 2
	Inj Volume : 1.000 µl

Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 1/5/2023 12:01:53 PM by SYSTEM
Method Info : Operator: Organic QC lab
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate
:0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron
particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole
MS Detector. No UV Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

