

File :Q:\GCMS-04 Minimal\DATA\090921\01 YX.D
Operator : Organic QC Lab
Acquired : 9 Sep 2021 13:25 using AcqMethod VOC-100SPLIT DECKER.M
Instrument : GCMS4
Sample Name: M-8015B-5031-14-TP-R1
Misc Info : Ethylene oxide in isooctane
Vial Number: 53



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

