


File : Q:\GCMS-06 Minimal\DATA\032921\27 -DC.D
Operator : Organic QC Lab
Acquired : 29 Mar 21 06:39 pm using AcqMethod CHICK_2019_13MIN.M
Instrument : GCMS 6
Sample Name: M-8015B-0531-20
Misc Info : N-Nitrosodi-n-butylamine in Water
Vial Number: 26

 **AccuStandard®**
Leader in Analytical Reference Standards
Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

