

File :Q:\GCMS-08 Minimal\DATA\20220105\ YX.D
Operator : Organic QC Lab
Acquired : 05 Jan 2022 14:22 using AcqMethod Hy_M502_S20.M
Instrument : GCMS 8
Sample Name: M-8015B-5031-21
Misc Info : Paraldehyde in water
Vial Number: 141



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

