

File : \\SVR-FS-01\Organic QC\QC Data\GCMS-08 Minimal\DATA\20231002  
... \06a- KM.D  
Operator : Organic QC Lab  
Instrument : GCMS 8  
Acquired : 02 Oct 2023 12:52 using AcqMethod Hy\_Pest\_15cm\_S10.M  
Sample Name: M-8081-SC  
Misc Info : Single-Colume Analyte Mix in 50:50 Hexane/Toluene



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

