

File :Q:\GCMS-08 Minimal\DATA\20221228\61- (JP)-.D  
Operator : Organic QC Lab  
Acquired : 29 Dec 2022 21:22 using AcqMethod Hy\_PAH\_15cm\_S100.M  
Instrument : GCMS 8  
Sample Name: M-8091-X1  
Misc Info : in iso:tol [1:1]  
Vial Number: 33



Column: DB-5, 30m, 0.25 ID, 0.25um  
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.  
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow  
Split 20:1, 1ul inj.; GC/MS; INJ 250c.

