

File : L:\QC Data\GCMS-06 Minimal\DATA\050522\20 - SAK.D
 Operator : Organic QC Lab
 Acquired : 05 May 2022 15:49 using AcqMethod CHICK_2019_S100.M
 Instrument : GCMS 6
 Sample Name: M-8120
 Misc Info : resample
 Vial Number: 77

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
 Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
 GC Parameters: Pulsed split 40psi 0.5m in, 10psi constant flow
 Split 2:1, 2µl inj.; GC/MS; Injector 270°C

