

=====

Acq. Operator : SYSTEM Seq. Line : 15

Sample Operator : SYSTEM

Acq. Instrument : LCMS-2 Location : P2-D-06

Injection Date : 12/5/2023 12:00:24 PM Inj : 1

Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\120523\120523 2023-12-05 07-35-33  
\Purity\_195-315.M

Last changed : 2/6/2023 2:02:18 PM by SYSTEM

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity\_195-315.M

Last changed : 2/6/2023 2:02:18 PM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min,the 100%  
Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse  
plus C-18, 2.1mmx50mm, 1.8 u particlue size. Agilent 1290 Infinity II LC/MS system  
, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

