

=====:
Acq. Operator : SYSTEM Seq. Line : 20
Acq. Instrument : LCMS-1 Location : P1-F5
Injection Date : 12/22/2021 1:04:32 PM Inj : 1
Inj Volume : 1.000 µl

Acq. Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 11/5/2021 2:37:17 PM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 12/22/2021 9:53:55 AM by SYSTEM
(modified after loading)
Method Info : Operator: Organic QC lab
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml
/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particle size.
Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV
Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

