

File : M:\GCMS-07 Minimal\DATA\06072101\20- ALR.D
 Operator : Organic QC department
 Acquired : 07 Jun 2021 16:40 using AcqMethod KERLA-2017.M
 Instrument : 5977B MSD
 Sample Name: M-8270-02-R1
 Misc Info :
 Vial Number: 30

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
 Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
 GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
 Split 20:1, 1µl inj.; GC/MS; Injector 270°C

