

File : \\SVR-FS-01\Organic QC\QC Data\GCMS-08 Minimal\DATA\20230404  
... \09- KM.D  
Operator : Organic QC Lab  
Instrument : GCMS 8  
Acquired : 04 Apr 2023 09:43 using AcqMethod Hy\_Pest\_15cm\_S100.M  
Sample Name: M-8270-06-ASL  
Misc Info : Method 8270B - PAH Mix in 50:50 DCM/Benzene



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

