

File :D:\MassHunter\GCMS\1\data\07181901\09- .D
 Operator : KD
 Acquired : 18 Jul 2019 10:04 using AcqMethod KERLA-2017_S50.M
 Instrument : 5977B MSD
 Sample Name: M-8270-14-ASL
 Misc Info : @2000ug/mL in Acetone
 Vial Number: 45

Column: DB-5, 30m, 0.25 ID, 0.25um
 Oven Program: 100c, 1m, 30c/min to 320c, final 2min.
 GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow
 Split 20:1, 1ul inj.; GC/MS; INJ 250c.

