

File :D:\MassHunter\GCMS\1\DATA\04062001\13- .D
Operator : Organic QC department
Acquired : 06 Apr 2020 11:35 using AcqMethod KERLA-2017.M
Instrument : 5977B MSD
Sample Name: M-8270-21
Misc Info in Acetone
Vial Number: 117

Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

