

File : Q:\GCMS-08\DATA\20181010\15-
 Operator : JJP / KM
 Acquired : 10 Oct 2018 19:33 using AcqMethod KERLA-2017-100uL draw speed.M
 Instrument : GCMS 8
 Sample Name: M-8270-WL-4X
 Misc Info : 8270 Working Level standard in MeCl2
 Vial Number: 89



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
 Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
 GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
 Split 20:1, 1µl inj.; GC/MS; Injector 270°C

