

File : \\SVR-FS-01\Organic QC\QC Data\GCMS-06 Minimal\DATA\120821\B  
... SB\32- KM.D  
Operator : Organic QC Lab  
Instrument : GCMS 6  
Acquired : 09 Dec 2021 01:28 using AcqMethod CHICK\_2019\_2S\_2uL.M  
Sample Name: M-8280A  
Misc Info : Dioxin Mix in Toluene



AccuStandard®

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

