

File : M:\GCMS-07 Minimal\DATA\09232101\51-ALR.D  
Operator : Organic QC department  
Acquired : 24 Sep 2021 11:07 using AcqMethod KERLA-2017\_Short\_15Min\_S100.M  
Instrument : 5977B MSD  
Sample Name : M-8310-SFE-IS-100X  
Misc Info : Biphenyl @ 20mg/mL in AcCN/THF 50/50 1:10 DIL AcCN  
Vial Number: 126



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

