



Acq. Operator : SYSTEM

Seq. Line : 2

Sample Operator : SYSTEM

Acq. Instrument : LCMS-2

Location : P2-E-01

Injection Date : 12/20/2023 12:42:22 PM

Inj : 2

Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\121923\121923 2023-12-20 12-24-32  
\DNPH.M

Last changed : 12/4/2019 12:27:07 PM by SYSTEM

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity\_195-315.M

Last changed : 2/6/2023 2:02:18 PM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

## Current Chromatogram(s)

