



=====

Acq. Operator : SYSTEM	Seq. Line : 2
Sample Operator : SYSTEM	
Acq. Instrument : LCMS-2	Location : P2-F-05
Injection Date : 8/6/2024 1:24:20 PM	Inj : 2
	Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\080624\080624 2024-08-06 13-06-32 \DNPH.M
Last changed : 12/4/2019 12:27:07 PM by SYSTEM
Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M
Last changed : 2/6/2023 2:02:18 PM by SYSTEM
Method Info : OPERATOR: Organic QC lab
General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system , Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

