

```
=====
Acq. Operator   : SYSTEM                      Seq. Line :    1
Acq. Instrument : LCMS-1                    Location  : P1-A1
Injection Date  : 5/14/2021 11:23:07 AM      Inj       :    1
                                           Inj Volume: 1.000 µl
```



```
Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 5/8/2020 11:20:44 AM by SYSTEM
Method Info : Operator: Oraganic QC lab
              Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
              Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate
              :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron
              particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole
              MS Detector. No UV Reference.
```

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

