



=====

Acq. Operator : SYSTEM	Seq. Line : 2
Sample Operator : SYSTEM	
Acq. Instrument : LCMS-2	Location : P2-C-01
Injection Date : 12/13/2023 8:01:55 AM	Inj : 2
	Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\121323\121323 2023-12-13 07-42-02  
          \Purity\_195-315.M  
Last changed : 2/6/2023 2:02:18 PM by SYSTEM  
Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity\_195-315.M  
Last changed : 2/6/2023 2:02:18 PM by SYSTEM  
Method Info : OPERATOR: Organic QC lab  
          General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%  
          Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse  
          plus C-18, 2.1mmx50mm, 1.8 µ particlue size. Agilent 1290 Infinity II LC/MS system  
          , Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

