

File : Q:\GCMS-08 Minimal\DATA\20190503\BSB\11- .D  
Operator : JP/KM  
Acquired : 03 May 2019 21:15 using AcqMethod KERLA-2Split-2uL.M  
Instrument : GCMS 8  
Sample Name: M-PNA-1  
Misc Info : PAH Mix in AcCN  
Vial Number: 91



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

