



AccuStandard®

File : Q:\GCMS-08\DATA\20180615\06
Operator : JJP / KM
Acquired : 15 Jun 2018 15:34 using AcqMethod KERLA-VOC-SV.M
Instrument : GCMS 8
Sample Name: NF-467-CUMENE
Misc Info : Cumene in DMF
Vial Number: 76

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

