



Sample Name: P-017R-WL-5X; 70-250 ng/ml in Iso

File: C:\CPData\DATA\20250318\JP(1).0000.RAW

Date (start): 3/18/2025 6:18:40 PM

Column: 110c, 0min, 30c/m to 250, 17c/min to 325, Final 5min=

Parameters: Splitless; 2ul; ECD: 325c; Inj: -250c

notes:

Method: Ni_30-250_17-325_ECD.MET

Setpoint File: Ni_30-250_17-325_ECD.M68

Calibration File:

Inst: GC #7

Oper: JJP

Vial = 2

4.84 Tetrachloro-meta-xylene (877-09-8)
7.17 Chlordane (Technical Mixture) (12789-03-6)
10.84 Decachlorobiphenyl (2051-24-3)

