



```

=====
Acq. Operator   : SYSTEM                      Seq. Line :    2
Acq. Instrument : HPLC only                  Location  : P1-A1
Injection Date  : 6/7/2019 9:32:37 AM       Inj       :    2
                                                Inj Volume: 1.000 µl

Acq. Method    : C:\Chem32\2\Data\06071901\06071901 2019-06-07 09-13-03\purity new.M
Last changed   : 6/7/2019 9:13:03 AM by SYSTEM
Analysis Method: D:\DATA\06051901\06051901 2019-06-05 13-34-24\Purity195_315.M (Sequence Method)
Last changed   : 6/10/2019 5:12:25 AM by SYSTEM
                (modified after loading)

Method Info    : Operator: GA
                Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml
                /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size.
                Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV
                Reference.
    
```

Sample Info : 2,4-D

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

