

File :D:\MassHunter\GCMS\1\data\07192023v2\16- ALR.D
Operator :
Acquired : 19 Jul 2023 15:58 using AcqMethod ALR 19 min 20 split FAST .M
Instrument : GCMS 9
Sample Name: P-024S
Misc Info : o,p'-DDD in MeOH
Vial Number: 16



Column: DB-5MS UI, 20m x 0.18mm ID x 0.18µm
Oven Program: 45°C, 13°C/min to 290°C (hold 5min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector temp 270°C

