

File :D:\MassHunter\GCMS\1\DATA\06272201\02- KD.D  
Operator : Organic QC department  
Acquired : 27 Jun 2022 07:20 using AcqMethod KERLA-2017\_Short\_15Min.M  
Instrument : 5977B MSD  
Sample Name: P-035S-CN-10X  
Misc Info :  
Vial Number: 37



Column: DB-5MS, 30m, 0.25 ID, 0.25 um  
Oven Program: 80c, 17c/min to 340c, 8min  
GC Parameters: Cons. Split, 12psi constant flow  
Split 20:1, 1uL inj.; GC/MS; INJ 270c

