

File : Q:\GCMS-08 Minimal\DATA\20200129\09-.D
Operator : Organic QC Lab
Acquired : 29 Jan 2020 13:29 using AcqMethod Hy_Pest_15cm_S20.M
Instrument : GCMS 8
Sample Name: P-057S
Misc Info : Kelthane in Methanol
Vial Number: 132



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

