


File :D:\MassHunter\GCMS\1\DATA\04062001\17- .D  
Operator : Organic QC department  
Acquired : 06 Apr 2020 13:34 using AcqMethod KERLA-2017\_Short\_15Min.M  
Instrument : 5977B MSD  
Sample Name: P-070S-10X  
Misc Info : Parathion in MeOH  
Vial Number: 119

 **AccuStandard**<sup>®</sup>  
Leader in Analytical Reference Standards  
Column: DB-5MS, 30m, 0.25 ID, 0.25 um  
Oven Program: 80c, 17c/min to 340c, 8min  
GC Parameters: Cons. Split, 12psi constant flow  
Split 20:1, 1uL inj.; GC/MS; INJ 270c

