

File :Q:\GCMS-07\DATA\06131801\13- : K.D  
Operator  
Acquired : 13 Jun 2018 14:30 using AcqMethod KERLA-2017\_Short\_15Min.M  
Instrument : 5977B MSD  
Sample Name: P-090S-CN-10X  
Misc Info : Tricyclazole @1000ug/mL in ACCN  
Vial Number: 43

Column: DB-5, 30m, 0.25 ID, 0.25um  
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.  
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow  
Split 20:1, 1ul inj.; GC/MS; INJ 250c.

