

File :Q:\GCMS-08 Minimal\DATA\20230103\26- =JP=.D
Operator : Organic QC Lab
Acquired : 03 Jan 23 08:18 pm using AcqMethod Hy_Pest_24cm_S50.M
Instrument : GCMS 8
Sample Name: P-0915-10X
Misc Info : Endosulfan I; in meoh
Vial Number: 31



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

