

=====

Acq. Operator : SYSTEM
Acq. Instrument : LCMS-1
Injection Date : 3/20/2025 8:09:00 PM

Seq. Line :
Location :
Inj :



AccuStandard®
Leader in Analytical Reference Standards

Inj Volume : 1.000 µl

Different Inj Volume from Sample Entry! Actual Inj Volume : 3.000 µl

Acq. Method : D:\ChemStation\1\Data\03202025\03202025 2025-03-20 09-51-44\Purity195_315.M

Last changed : 1/22/2025 1:03:02 PM by SYSTEM

Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M

Last changed : 2/6/2023 2:02:18 PM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

