



=====
Acq. Operator : SYSTEM Seq. Line : 23
Acq. Instrument : LCMS-1 Location : P2-C5
Injection Date : 6/18/2018 10:54:37 PM Inj : 1
Inj Volume : 1.000 µl

Acq. Method : D:\DATA\06181801\060181801 2018-06-18 16-04-05\Purity195_315.M
Last changed : 6/18/2018 4:04:06 PM by SYSTEM
Analysis Method : C:\Chem32\1\Methods\Purity195_315.M
Last changed : 6/18/2018 9:35:46 AM by SYSTEM
(modified after loading)

Method Info : Operator: GA
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.
3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule
size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector.
No UV Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

