

Data File : D:\DATA\03282022\06-KD.D

Sample Name : P-1060S

=====
Acq. Operator : SYSTEM Seq. Line : 6
Acq. Instrument : LCMS-1 Location : P1-B9
Injection Date : 3/28/2022 8:55:54 AM Inj : 1
 Inj Volume : 1.000 µl

Acq. Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 11/5/2021 2:37:17 PM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 3/16/2022 2:44:36 PM by SYSTEM
(modified after loading)
Method Info : Operator: Organic QC lab
 Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
 Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml
 /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particle size.
 Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV
 Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

