Print of window 38: Current Chromatogram(s)
Data File : D:\DATA\07101801\07101801 2018-07-1016-10-25\
Sample Name : P-1066S-CN-10X Oxaziclomefone

AccuStandard®

Acq. Operator	: SYSTEM	Seq. Line :	23
Acq. Instrument	: LCMS-1	Location :	P1-C5
Injection Date	: 7/10/2018 11:00:55 PM	Inj :	1
		Inj Volume :	1.000 µl

Acq. Method : D:\DATA\07101801\07101801 2018-07-10 16-10-25\Purity195_315.M
Last changed : 7/10/2018 4:10:25 PM by SYSTEM
Analysis Method : C:\Chem32\1\Methods\Purity195_315.M
Last changed : 7/11/2018 9:56:54 AM by SYSTEM
(modified after loading)
Method Info : Operator: GA
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)