

=====

Acq. Operator	: SYSTEM	Seq. Line	: 4
Acq. Instrument	: LCMS-1	Location	: P2-E7
Injection Date	: 7/23/2018 2:53:04 PM	Inj	: 1
		Inj Volume	: 1.000 µl

Acq. Method : C:\Chem32\1\Data\07231805 2018-07-23 13-56-29\PURITY195_315.M
Last changed : 7/23/2018 1:56:29 PM by SYSTEM
Analysis Method : C:\Chem32\1\Data\07231802\07231802 2018-07-23 11-29-02\Purity195_315.M (Sequence Method)
Last changed : 7/23/2018 11:29:02 AM by SYSTEM
Method Info : Operator: GA
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV Reference.

Current Chromatogram(s)

