

```

=====
Acq. Operator   : SYSTEM                      Seq. Line :    2
Acq. Instrument : LCMS-1                    Location  :   P1-B1
Injection Date  : 3/19/2020 6:25:10 AM      Inj       :    1
                                           Inj Volume: 1.000 µl
  
```

```

Acq. Method    : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed   : 10/18/2019 4:18:14 PM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed   : 3/10/2020 2:22:01 PM by SYSTEM
                (modified after loading)
  
```

```

Method Info    : Operator: GA
                Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate
                :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron
                particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole
                MS Detector. No UV Reference.
  
```

```
Sample Info    : Ethaboxam in ACCN
```

```
Additional Info : Peak(s) manually integrated
```

Current Chromatogram(s)

