

```

=====
Acq. Operator   : SYSTEM                               Seq. Line :   11
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2                               Location  :   P1-F-03
Injection Date  : 6/25/2020 7:30:44 PM                 Inj       :    1
                                                    Inj Volume: 1.000 µl
Acq. Method     : C:\Users\Public\Documents\ChemStation\1\Data\06252001\06242001 2020-06-25 16-20-22
                  \Purity_195-315.M
Last changed    : 5/8/2020 10:12:19 AM by SYSTEM
Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M
Last changed    : 6/24/2020 12:11:23 PM by SYSTEM
                  (modified after loading)
Method Info     : OPERATOR: Organic QC lab
                  General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%
                  Acetonitile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse
                  plus C-18, 2.1mmx50mm, 1.8 µ particlue size. Agilent 1290 Infinity II LC/MS system
                  , Agilent 6125 quadrupole MS detector.
    
```

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

