



=====
 Acq. Operator : SYSTEM Seq. Line : 10
 Sample Operator : SYSTEM
 Acq. Instrument : LCMS-2 Location : P2-D-04
 Injection Date : 8/21/2023 9:15:24 AM Inj : 1
 Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\082123\082123101 2023-08-21 06-24-54\Purity_195-315.M
 Last changed : 2/6/2023 2:02:18 PM by SYSTEM
 Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M
 Last changed : 8/16/2023 11:04:06 AM by SYSTEM
 (modified after loading)

Method Info : OPERATOR: Organic QC lab
 General gradient method starting @100% water for 1 min @0.3 ml/min,the 100% Acetonitile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 u particlue size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

