

=====
Acq. Operator : SYSTEM Seq. Line : 8
Acq. Instrument : LCMS-1 Location : P1-D5
Injection Date : 4/26/2021 10:06:37 AM Inj : 1
 Inj Volume : 1.000 µl

Acq. Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 5/8/2020 11:20:44 AM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 4/20/2021 4:06:14 PM by SYSTEM

Method Info : Operator: Oraganic QC lab
 Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
 Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml
 /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size.
 Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV
 Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

