



=====

Acq. Operator	: SYSTEM	Seq. Line	: 5
Acq. Instrument	: LCMS-1	Location	: P1-A5
Injection Date	: 7/5/2018 4:55:01 PM	Inj	: 1
		Inj Volume	: 1.000 µl

Acq. Method : D:\DATA\07051801\07051801 2018-07-05 15-39-50\Purity195_315.M

Last changed : 7/5/2018 3:39:50 PM by SYSTEM

Analysis Method : C:\Chem32\1\Methods\Purity195_315.M

Last changed : 7/5/2018 12:29:09 PM by SYSTEM

(modified after loading)

Method Info : Operator: GA

Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

