

File :D:\MassHunter\GCMS\1\data\20220616\02 - SAK.D  
Operator : Organic QC Lab  
Acquired : 16 Jun 2022 08:16 using AcqMethod Hy\_Pest\_24cm\_S10.M  
Instrument : GCMS 8  
Sample Name: P-122S  
Misc Info :  
Vial Number: 119



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

