



```

=====
Acq. Operator   : SYSTEM                      Seq. Line :   21
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2                      Location  :   P2-D-08
Injection Date  : 11/23/2021 2:56:48 PM      Inj       :    1
                                                Inj Volume: 1.000 µl
    
```

```

Acq. Method    : C:\Users\Public\Documents\ChemStation\1\Data\11232101\11232101 2021-11-23 08-37-45
                 \Purity_195-315.M
    
```

```

Last changed   : 4/29/2021 10:41:07 AM by SYSTEM
    
```

```

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M
    
```

```

Last changed   : 4/29/2021 10:41:07 AM by SYSTEM
    
```

```

Method Info    : OPERATOR: Organic QC lab
    
```

```

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%
Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse
plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system
, Agilent 6125 quadrupole MS detector.
    
```

```

Additional Info : Peak(s) manually integrated
    
```

Current Chromatogram(s)

