



=====  
Acq. Operator : SYSTEM Seq. Line : 12  
Acq. Instrument : LCMS-1 Location : P1-C6  
Injection Date : 8/28/2023 10:13:54 AM Inj : 1  
Inj Volume : 1.000 µl

Method : C:\CHEM32\1\METHODS\Purity195\_315.M  
Last changed : 1/5/2023 12:01:53 PM by SYSTEM  
Method Info : Operator: Organic QC lab  
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %  
Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate  
:0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron  
particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole  
MS Detector. No UV Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

