



=====  
Acq. Operator : SYSTEM Seq. Line : 11  
Sample Operator : SYSTEM  
Acq. Instrument : LCMS-2 Location : P2-E-07  
Injection Date : 11/22/2024 1:33:24 PM Inj : 1  
Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\112224\112124 2024-11-22 10-23-50\Purity\_195-315.M  
Last changed : 2/6/2023 2:02:18 PM by SYSTEM  
Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity\_195-315.M  
Last changed : 2/6/2023 2:02:18 PM by SYSTEM  
Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

